



New Hampshire Fish and Game Department  
Powder Mill Fish Hatchery  
288 Merrymeeting Road  
New Durham, NH 03855  
Telephone: 603/859-2041

July 19, 2018

Norma Mason  
US Environmental Protection Agency  
Water Technical Unit - SEW  
PO Box 8127  
Boston, MA 02114-8127

**RE: Federal Permit Number: NH0000710  
Discharge 002**

Dear Ms. Mason:

I enclosed the September 2014 Discharge Monitoring Reports for the Powder Mill Fish Hatchery for Discharge 002.

Upstream water sample for pH was performed when pH was less then 6.5. The pH Less then 0.5 S.U. difference into receiving waters so compliance was met

pH analysis was not conducted the week of 09/15 to 09/19 due to the facility meter being sent out for annual calibration.

Facilities spare meter was not found in time to perform test. Spare meter will be kept at this location in the future to prevent this from happening again.

The weekly tests are performed by the Fish Culturists at Powder Mill Fish Hatchery.

Thomas W. Givetz

---

Thomas W. Givetz

Superintendent  
Powder Mill State Fish Hatchery

